## Technical Specifications

	75	

TG 309 Libra®					
	Classic	Select	Supreme		
Temperature range	(10°C) RT to 1025°C	(10°C) RT to 1025°C/1100°C	(10°C) RT to 1100°C		
Heating rate	0.001 K/min to 200 K/min	0.001 K/min to 200 K/min	0.001 K/min to 200 K/min		
Balance resolution	50 ng	20 ng	10 ng		
Cooling time <sup>1</sup>	In nitrogen: $\approx$ 12 min from 1100°C to 100°C In helium: $\approx$ 5 min from 1100°C to 100°C, $\approx$ 10 min to 25°C				
AutoVac	Automatic evacuation and refilling of purge gas; optionally available if MFC is selected				
Temperature resolution	0.001 K	0.001 K	0.001 K		
Temperature accuracy <sup>2</sup>	± 0.3 K (after calibration by c- <i>DTA</i> ®, indium)				
Temperature calibration	c-DTA®, also for detection of endo- and exothermal effects; Curie standards				
Temperature stability <sup>3</sup>	Peak-to-peak: 0.03 K RMS: 0.005 K				
Temperature precision <sup>4</sup>		0.15 K			
Vacuum-tightness	1 mbar	<< 10 <sup>-1</sup> mbar	<< 10 <sup>-1</sup> mbar		
Integrated 4-fold MFC					
Integrated 3-fold MFC			-		
Evolved Gas Analysis	-				
192+12-position ASC	-				
20-position ASC		_	-		
Piercing device	-				
Color touch display					
Unlimited warranty⁵					

1 21°C chiller temperature, 200 ml/min He (purge + protective gas); the maximum

temperature of the TGA system depends on the He gas flow: at 200 ml/min, T<sub>max</sub> is 1020°C.

2 Maximum deviation between measured and literature value (indium)

3 Measurement at different isothermal temperatures

4 Standard deviation based on 10 measurements

5 In connection with maintenance contract



included

optional

Traunstraße 21, A-2120 Wolkersdorf T:+43 2245 6725 F:+43 2245 559633 office@prager-elektronik.at www.prager-elektronik.at